

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130571</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>STEVE DUNN</u>	ID# <u>4106</u> ✓ Date <u>3-30-11</u> ✓
A Agency <u>APD</u>	Phone # <u>786-8900</u>
Instrument Location <u>SEELY SUB (SPENARD)</u>	
B Alco S/N <u>2V</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>200 PSI</u>	
Alco Test Values <u>.079</u> ✓ <u>.080</u> ✓	
1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2011 aw

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

S. Dunn #1361  
ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

MARCH 30, 2011

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

V

O.L. #: V

DEPT/AGENCY: ANC1

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 2V

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.38 in  
ALCO TARGET .078 11:56  
BLANK TEST .000 11:57  
INTERNAL STANDARD VERIFIED 11:57  
ALCO TV 29.38 in .079 11:57  
BLANK TEST .000 11:58  
SUBJECT SAMPLE .000 11:59  
BLANK TEST .000 11:59  
ALCO TV 29.38 in .000 12:00  
BLANK TEST .000 12:00

## DIAGNOSTIC CHECK

S. Dunn #1361  
ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

MARCH 30, 2011

TIME 11:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/05

HEATERS

SAMPLE CHAMBER: 53c

BREATH TUBE: 42c

BAROMETER: 29.38 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~"